

**MAA519Hu26** 

Monoclonal Antibody to Apolipoprotein A1 (APOA1)

Organism Species: Homo sapiens (Human)

Instruction manual

FOR RESEARCH USE ONLY NOT FOR USE IN CLINICAL DIAGNOSTIC PROCEDURES

12th Edition (Revised in Aug, 2016)



# [PROPERTIES]

**Source:** Monoclonal antibody preparation

Host: Mouse

Antibody isotype: IgG2b Kappa

**Purification:** Protein A + Protein G affinity chromatography

Clone number: C16

Traits: Liquid

Concentration: 1mg/ml

**UOM:** 100µl

Cross Reactivity: Rat

Applications: WB; IHC; ICC; IP.

### [ IMMUNOGEN ]

Immunogen: Recombinant APOA1 (Gln122~Gln267) expressed in E.coli

Accession No.: RPA519Hu01

#### [ APPLICATIONS ]

Western blotting: 0.5-2µg/mL;

Immunohistochemistry: 5-20µg/mL;

Immunocytochemistry: 5-20µg/mL;

Optimal working dilutions must be determined by end user.

#### [FORMULATION]

Form & Buffer: Supplied as solution form in 0.01M PBS, pH7.4, containing 0.05% Proclin-300,

50% glycerol.

#### [ STORAGE AND STABILITY ]

Storage: Avoid repeated freeze/thaw cycles.

Store at 4°C for frequent use.

Aliquot and store at -20°C for 24 months.

Stability Test: The thermal stability is described by the loss rate. The loss rate was determined

# Coud-Clone Corp.

by accelerated thermal degradation test, that is, incubate the protein at 37°C for 48h, and no obvious degradation and precipitation were observed. The loss rate is less than 5% within the expiration date under appropriate storage condition.

# [ IDENTIFICATION ]




Figure. Western Blot; Sample: Recombinant APOA1, Human.

Western Blot; Sample: Human Serum; Western Blot; Sample: Rat Cerebrum Primary Ab: 2µg/ml Mouse Anti-Human lysate;

APOA1 Antibody
Second Ab: 0.2µg/mL HRP-Linked
Caprine Anti-Mouse IgG Polyclonal
Antibody
(Catalog: SAA544Mu19)

lysate;
Primary Ab: 2µg/ml Mouse Anti-Human
APOA1 Antibody
Second Ab: 0.2µg/mL HRP-Linked
Caprine Anti-Mouse IgG Polyclonal
Antibody

(Catalog: SAA544Mu19)

# [ IMPORTANT NOTE ]

The kit is designed for research use only, we will not be responsible for any issue if the kit was used in clinical diagnostic or any other procedures.