

**PAA817Rb52** 

Polyclonal Antibody to Haptoglobin (Hpt)

Organism Species: Oryctolagus cuniculus (Rabbit)

Instruction manual

FOR RESEARCH USE ONLY NOT FOR USE IN CLINICAL DIAGNOSTIC PROCEDURES

13th Edition (Revised in Aug, 2023)



## [PROPERTIES]

Source: Polyclonal antibody preparation

Host: Guinea pig

**Purification:** Antigen-specific affinity chromatography followed by Protein A affinity

chromatography

Traits: Liquid

Concentration: 0.5mg/ml

**UOM:** 200µl

Cross Reactivity: N/A

Applications: WB; IHC.

### [IMMUNOGEN]

Immunogen: Recombinant Hpt (Val177~Lys342) expressed in E.coli

Accession No.: RPA817Rb02

### [ APPLICATIONS ]

Western blotting: 0.01-2µg/mL

Immunohistochemistry: 5-20µg/mL

Immunocytochemistry: 5-20µg/mL

Optimal working dilutions must be determined by end user.

#### [FORMULATION]

**Form & Buffer:** Supplied as solution form in 0.01M PBS, pH7.4, containing 0.05% Proclin-300, 50% glycerol.

#### [ STORAGE AND STABILITY ]

**Storage:** Avoid repeated freeze/thaw cycles.

Store at 4°C for frequent use.

Aliquot and store at -20°C for 24 months.

**Stability Test:** The thermal stability is described by the loss rate. The loss rate was determined by accelerated thermal degradation test, that is, incubate the protein at 37°C for 48h, and no



obvious degradation and precipitation were observed. The loss rate is less than 5% within the expiration date under appropriate storage condition.

## [ IDENTIFICATION ]





Figure. Western Blot; Sample: Recombinant Hpt, Rabbit.



Western Blot; Sample: Rabbit Serum
Primary Ab: 1µg/ml Cavia Anti-Rabbit
Hpt Antibody
Second Ab: 0.2µg/mL HRP-Linked
Rabbit Anti-Cavia IgG Polyclonal
Antibody

(Catalog: SAA544Gu09)



DAB staining on IHC-P; Samples:
Rabbit Liver Tissue; Primary Ab:
10µg/ml Cavia Anti-Rabbit Hpt Antibody
Second Ab: 2µg/mL HRP-Linked Rabbit
Anti-Cavia IgG Polyclonal Antibody
(Catalog: SAA544Gu09)

# [ IMPORTANT NOTE ]

The kit is designed for research use only, we will not be responsible for any issue if the kit was used in clinical diagnostic or any other procedures.