

PAA899Si01

Polyclonal Antibody to Osteopontin (OPN)

Organism Species: Rhesus monkey (Simian)

Instruction manual

FOR RESEARCH USE ONLY

NOT FOR USE IN CLINICAL DIAGNOSTIC PROCEDURES

12th Edition (Revised in Aug, 2016)

Cond-Clone Corp.

[PROPERTIES]

Source: Polyclonal antibody preparation

Host: Rabbit

Purification: Antigen-specific affinity chromatography followed by Protein A affinity

chromatography

Traits: Liquid

Concentration: 0.84mg/ml

UOM: 100µg(120µl)

Cross Reactivity: Human; Pig

Applications: WB; IHC; ICC; IP.

[<u>IMMUNOGEN</u>]

Immunogen: Recombinant OPN (Pro18~Asn314 (Accession # F6V2X3)) expressed in E.coli

Accession No.: RPA899Si01

[APPLICATIONS]

Western blotting: 0.5-3µg/mL

Immunohistochemistry: 5-30µg/mL

Immunocytochemistry: 5-30µg/mL

Optimal working dilutions must be determined by end user.

[FORMULATION]

Form & Buffer: Supplied as solution form in 0.01M PBS, pH7.4, containing 0.05% Proclin-300, 50% glycerol.

[STORAGE AND STABILITY]

Storage: Avoid repeated freeze/thaw cycles.

Store at 4°C for frequent use.

Aliquot and store at -20°C for 24 months.

Stability Test: The thermal stability is described by the loss rate. The loss rate was determined by accelerated thermal degradation test, that is, incubate the protein at 37°C for 48h, and no

Cloud-Clone Corp.

obvious degradation and precipitation were observed. The loss rate is less than 5% within the expiration date under appropriate storage condition.








Figure. Western Blot; Sample:

Recombinant OPN, Simian.



lysate;

OPN Antibody



DAB staining on IHC-P;

Samples: Human Glioma Tissue; Western Blot; Sample: Porcine Kidney Primary Ab: 10µg/ml Rabbit Anti-Simian **OPN** Antibody Primary Ab: 3µg/ml Rabbit Anti-Simian Second Ab: 2µg/mL HRP-Linked Caprine Anti-Rabbit IgG Polyclonal Second Ab: 0.2µg/mL HRP-Linked Antibody Caprine Anti-Rabbit IgG Polyclonal (Catalog: SAA544Rb19)

Antibody (Catalog: SAA544Rb19)

[<u>IMPORTANT NOTE</u>]

The kit is designed for research use only, we will not be responsible for any issue if the kit was used in clinical diagnostic or any other procedures.