

RPC919Hu03 50µg Recombinant Platelet Derived Growth Factor D (PDGFD) Organism Species: *Homo sapiens (Human)* Instruction manual

FOR RESEARCH USE ONLY

NOT FOR USE IN CLINICAL DIAGNOSTIC PROCEDURES

12th Edition (Revised in Aug, 2016)

Coud-Clone Corp.

[PROPERTIES]

Source: Prokaryotic expression

Host: E.coli

Residues: Arg19~Arg370

Tags: N-terminal His and GST Tag

Subcellular Location: Secreted

Purity: > 90%

Traits: Freeze-dried powder

Buffer formulation: PBS, pH7.4, containing 0.01% SKL, 5% Trehalose.

Original Concentration: 200µg/mL

Applications: Positive Control; Immunogen; SDS-PAGE; WB.

(May be suitable for use in other assays to be determined by the end user.)

Predicted isoelectric point: 8.4

Predicted Molecular Mass: 70.7kDa

Accurate Molecular Mass: 72kDa as determined by SDS-PAGE reducing conditions.

[<u>USAGE</u>]

Reconstitute in 10mM PBS (pH7.4) to a concentration of 0.1-1.0 mg/mL. Do not vortex.

[STORAGE AND STABILITY]

Storage: Avoid repeated freeze/thaw cycles.

Store at 2-8°C for one month.

Aliquot and store at -80°C for 12 months.

Stability Test: The thermal stability is described by the loss rate. The loss rate was determined by accelerated thermal degradation test, that is, incubate the protein at 37°C for 48h, and no obvious degradation and precipitation were observed. The loss rate is less than 5% within the expiration date under appropriate storage condition.

[SEQUENCE]

Cond-Clone Corp.

RD TSATPQSASI KALRNANLRR DESNHLTDLY RRDETIQVKG NGYVQSPRFP NSYPRNLLLT WRLHSQENTR IQLVFDNQFG LEEAENDICR YDFVEVEDIS ETSTIIRGRW CGHKEVPPRI KSRTNQIKIT FKSDDYFVAK PGFKIYYSLL EDFQPAAASE TNWESVTSSI SGVSYNSPSV TDPTLIADAL DKKIAEFDTV EDLLKYFNPE SWQEDLENMY LDTPRYRGRS YHDRKSKVDL DRLNDDAKRY SCTPRNYSVN IREELKLANV VFFPRCLLVQ RCGGNCGCGT VNWRSCTCNS GKTVKKYHEV LQFEPGHIKR RGRAKTMALV DIQLDHHERC DCICSSRPPR

[IDENTIFICATION]




Figure . Gene Sequencing (extract)




Figure. SDS-PAGE

[IMPORTANT NOTE]

The kit is designed for research use only, we will not be responsible for any issue if the kit was used in clinical diagnostic or any other procedures.