

RPC512Hu01 10µg

Recombinant Prostate-specific Membrane Antigen (PMSA)

Organism Species: Homo sapiens (Human)

Instruction manual

FOR RESEARCH USE ONLY
NOT FOR USE IN CLINICAL DIAGNOSTIC PROCEDURES

12th Edition (Revised in Aug, 2016)



# [PROPERTIES]

**Source:** Prokaryotic expression

Host: E.coli

Residues: Ala274~Leu587

Tags: N-terminal His Tag

**Subcellular Location:** Membrane

**Purity:** > 95%

Traits: Freeze-dried powder

**Buffer formulation:** 20mM Tris, 150mM NaCl, pH8.0, containing 0.01% skl, 5%Trehalose.

Original Concentration: 350µg/mL

Applications: Positive Control; Immunogen; SDS-PAGE; WB.

(May be suitable for use in other assays to be determined by the end user.)

Predicted isoelectric point: 7.3

Predicted Molecular Mass: 39.1kDa

**Accurate Molecular Mass:** 39kDa as determined by SDS-PAGE reducing conditions.

#### [USAGE]

Reconstitute in 20mM Tris, 150mM NaCl (pH8.0) to a concentration of 0.1-1.0 mg/mL. Do not vortex.

#### [ STORAGE AND STABILITY ]

Storage: Avoid repeated freeze/thaw cycles.

Store at 2-8°C for one month.

Aliquot and store at -80°C for 12 months.

**Stability Test:** The thermal stability is described by the loss rate. The loss rate was determined by accelerated thermal degradation test, that is, incubate the protein at 37°C for 48h, and no obvious degradation and precipitation were observed. The loss rate is less than 5% within the expiration date under appropriate storage condition.

### [ SEQUENCE ]

```
ANEYAYR RGIAEAVGLP SIPVHPIGYY
DAQKLLEKMG GSAPPDSSWR GSLKVPYNVG PGFTGNFSTQ KVKMHIHSTN
EVTRIYNVIG TLRGAVEPDR YVILGGHRDS WVFGGIDPQS GAAVVHEIVR
SFGTLKKEGW RPRRTILFAS WDAEEFGLLG STEWAEENSR LLQERGVAYI
NADSSIEGNY TLRVDCTPLM YSLVHNLTKE LKSPDEGFEG KSLYESWTKK
SPSPEFSGMP RISKLGSGND FEVFFQRLGI ASGRARYTKN WETNKFSGYP
LYHSVYETYE LVEKFYDPMF KYHLTVAQVR GGMVFEL
```

### [ IDENTIFICATION ]




Figure . Gene Sequencing (extract)




Figure. SDS-PAGE

# [ IMPORTANT NOTE ]

The kit is designed for research use only, we will not be responsible for any issue if the kit was used in clinical diagnostic or any other procedures.